

Plate Type	Well Volume (µL)	Well Shape	Well Depth (mm)	Well Diameter (or Side for Square Well) top/bottom (mm)	Plate Length (mm)	Plate Width (mm)	Plate Height (mm)	A1 Row Offset -X (mm)	A1 Column Offset -Y (mm)	Well center to well center spacing (mm)	Flange or skirt height (mm)	Stack Height with Lid	Stack Height w/o Lid	Well Bottom Elevation (mm)	Well Bottom Thickness (microns)	Well Bottom Area (mm2)	Perimeter Evaporation Barrier
384 IQ LV (Imager Quality / Low Volume) 100 Micron Clear Film Bottom Part Numbers ABA1-xxxxx	48.94	RD	7.90	3.07/2.52	127.76	85.50	10.40	8.99	12.12	4.50	2.50	12.20	10.00	2.50	100	4.99	Yes
384 IQ LV (Imager Quality / Low Volume) 200 Micron Clear Film Bottom Part Numbers ABA2-xxxxx	48.94	RD	7.90	3.07/2.52	127.76	85.50	10.40	8.99	12.12	4.50	2.50	12.20	10.00	2.50	200	4.99	Yes
384 IQ LV (Low Volume) Solid Bottom Part Numbers ABA0-xxxxx, AWA0-xxxxx	48.94	RD	7.90	3.07/2.52	127.76	85.50	10.40	8.99	12.12	4.50	2.50	12.20	10.00	2.50	900	4.99	Yes
384 IQ EB (Imager Quality/ EvapBarrier) 100 Micron Clear Film Bottom Part Numbers ABC1-xxxxx	126.39	SQ	12.35	3.60/2.74	127.76	85.50	14.35	8.99	12.12	4.50	4.00	15.80	14.00	2.00	100	7.51	Yes
384 IQ EB (Imager Quality/ EvapBarrier) 200 Micron Clear Film Bottom Part Numbers ABC2-xxxxx	126.39	SQ	12.35	3.60/2.74	127.76	85.50	14.35	8.99	12.12	4.50	4.00	15.80	14.00	2.00	200	7.51	Yes
384 IQ EB (Imager Quality/ EvapBarrier) Solid Bottom Part Numbers ABC0-xxxxx, AWC0-xxxxx	126.39	SQ	12.35	3.60/2.74	127.76	85.50	14.35	8.99	12.12	4.50	4.00	15.80	14.00	2.00	900	7.51	Yes
384 IQ (Imager Quality) 100 Micron Clear Film Bottom Part Numbers ABB1-xxxxx	126.39	SQ	12.35	3.60/2.74	127.76	85.50	14.35	8.99	12.12	4.50	4.00	15.80	14.00	2.00	100	7.51	No
384 IQ (Imager Quality) 200 Micron Clear Film Bottom Part Numbers ABB2-xxxxx	126.39	SQ	12.35	3.60/2.74	127.76	85.50	14.35	8.99	12.12	4.50	4.00	15.80	14.00	2.00	200	7.51	No
384 IQ (Imager Quality) Solid Bottom Part Numbers ABB0-xxxxx, AWB0-xxxxx	126.39	SQ	12.35	3.60/2.74	127.76	85.50	14.35	8.99	12.12	4.50	4.00	15.80	14.00	2.00	900	7.51	No
384 IQ-ULB (Imager Quality) 200 Micron Clear Film Bottom Part Numbers ABD2-xxxxx	126.39	SQ	13.44	3.60/2.74	127.76	85.50	14.35	8.99	12.12	4.50	4.00	15.80	14.00	0.69	200	7.51	No
384 IQ-ULB-EB (Imager Quality/Evap Barrier) 200 Micron Clear Film Bottom Part Numbers ABE2-xxxxx	126.39	SQ	13.44	3.60/2.74	127.76	85.50	14.35	8.99	12.12	4.50	4.00	15.80	14.00	0.69	200	7.51	Yes